



Ambulanc (Shenzhen) Tech Co. Ltd.

## Emergency transport and Critical Cure Ventilator of T6 Certificate of Analysis

Product name: Emergency transport and Critical Cure Ventilator

Model:T6

Batch No.: T62110019

Test items	requirements				Test results
Appearance	There are no scratch, sags and crests, cracks, smudges, etc. on the appearance with uniform color. All structural connections are reliable, and the gaps between the connecting joints are uniform.				<input checked="" type="checkbox"/> Qualified <input type="checkbox"/> Unqualified
label	The label is pasted flat and firm and the silkscreen is clear, with correct contents and direction and without skewing.				<input checked="" type="checkbox"/> Qualified <input type="checkbox"/> Unqualified
Functional	Shuttle: The shuttle can control movement, selection, confirmation, parameter adjustment and confirmation after startup. Button: press the host button,the host implements the corresponding button function. Touch screen: Touch function is normal.				<input checked="" type="checkbox"/> Qualified <input type="checkbox"/> Unqualified
Control Parameter precision	Respiratory rate	1	0 ~ 2	1	<input checked="" type="checkbox"/> Qualified <input type="checkbox"/> Unqualified
	1 ~ 100:±1	60	59 ~ 61	60.4	
	Above 100: ±5% of	100	95 ~ 105	100.5	
	Set value (unit: bpm)	150	142.5~157.5	151.1	
	Inspiration time	0.2	0.1~0.3	0.22	<input checked="" type="checkbox"/> Qualified <input type="checkbox"/> Unqualified
	±0.1 or ±10% of set	1	0.9~1.1	1.03	
	value, whichever is	2	1.8~2.2	2.01	
	greater	10	9~11	10.08	

(unit: s)				
Inspiration/expiration ratio 2:1 ~ 1:4: ±10% of set value Other: ±15% of set value	4:1	4.6:1 ~ 3.4: 1	3.97:1	<input checked="" type="checkbox"/> Qualified <input type="checkbox"/> Unqualified
	1:1	0.9:1 ~ 1:0.9	1.05:1	
	1:10	1:8.5 ~ 1:11.5	1:9.11	
Tidal volume ± (1.5ml + 15% of set value) (Neonate)  ± (10ml + 10% of set value) (Adult/Pediatric) (unit: ml)	2	0.2 ~ 3.8	1.7	<input checked="" type="checkbox"/> Qualified <input type="checkbox"/> Unqualified
	10	7 ~ 13	10.1	
	20	15.5 ~ 24.5	18.0	
	100	80 ~ 120	103.3	
	300	260 ~ 340	300.1	
	1000	890 ~ 1110	999	
	2000	1790 ~ 2210	1929	
FiO <sub>2</sub> ± (3vol.%+1% of set value) (unit: %)	21	18~24	20.8	<input checked="" type="checkbox"/> Qualified <input type="checkbox"/> Unqualified
	40	37~43	40.8	
	60	57~63	61.2	
	70	67~73	70.9	
	100	97~100	100.6	
Inspiratory pressure ± (0.9 cmH <sub>2</sub> O ±10% of set value) (unit: cmH <sub>2</sub> O)	1	0~2	0.9	<input checked="" type="checkbox"/> Qualified <input type="checkbox"/> Unqualified
	40	35.1~44.9	40.75	
	60	53.1~66.9	60.76	
	90	80.1 ~ 99.9	91.01	
CPAP/PEEP ± (0.9 cmH <sub>2</sub> O ±10% of set value) (unit: cmH <sub>2</sub> O)	0	0~0.9	0.9	<input checked="" type="checkbox"/> Qualified <input type="checkbox"/> Unqualified
	10	8.1~11.9	10.15	
	20	18.1 ~ 21.9	20.06	
	40	36~44	40.88	
Oxygen flow rate ±2 or ±15 of set value, whichever is greater	2	0~4	1.99	<input checked="" type="checkbox"/> Qualified <input type="checkbox"/> Unqualified
	20	17~23	19.4	
	25	35.5~34.5	24.9	
	65	55.25~74.75	61.7	

	(unit: L/min)				
Monitoring Parameter precision	Respiratory frequenc ±2bpm or ±5% of reading value, whichever is greater (unit: bpm)	1	0~4	1	<input checked="" type="checkbox"/> Qualified <input type="checkbox"/> Unqualified
		60.2	58~62	60.2	
		101	95~105	100.5	
		151	142.5~157.5	151.1	
		250	237.5~262.5	250	
	Tidal Volume 0 ~ 100: ± (2+15% of reading value) 100 ~ 300: ± (3+15% of reading value) Above 300: ±15% of reading value (unit: ml)	1	0 ~ 3.61	1.9	<input checked="" type="checkbox"/> Qualified <input type="checkbox"/> Unqualified
		9	4.9 ~ 11.3	8.1	
		19	12.9 ~ 22.1	19.7	
		95	71.7 ~ 104.1	104.9	
		982	849.5 ~ 1148.5	999	
		2048	1634.5 ~ 2211.5	1942	
	Oxygen concentration ± (2.5vol.#+2.5% of reading value) (unit: %)	21	17.8 ~ 23.8	20.8	<input checked="" type="checkbox"/> Qualified <input type="checkbox"/> Unqualified
		40	37.3 ~ 44.3	40.8	
		60	57.2 ~ 65.2	61.2	
		70	66.6 ~ 75.2	70.9	
		99	98.1 ~ 103.1	100.6	
	Airway pressure ± (2 ±4% of reading value) (unit:: cmH20)	0.7	0 ~ 3.1	0.8	<input checked="" type="checkbox"/> Qualified <input type="checkbox"/> Unqualified
		40.7	37.7 ~ 45.0	41.8	
		60.7	56.7 ~ 65.6	61.2	
		90.4	86 ~ 97.4	90.2	
		103	97.2 ~ 109.5	103.3	
	CO <sub>2</sub> concentration 0 ~ 40: ±2 41 ~ 70: ±5% of reading value	38 (5%)	36~40	38.6	<input checked="" type="checkbox"/> Qualified <input type="checkbox"/> Unqualified
		60.8 (8%)	57.4~64.1	62.3	
		149.7 (19.7%)	135~165	152.4	

	71 ~ 100: ± 8% of reading value 101 ~ 150: ± 10% of reading value (unit: mmHg)				
Alarm Function	Air source pressure alarm, Airway pressure alarm, Tidal volume alarm, Minute ventilation alarm, Respiratory frequency alarm, Oxygen concentration alarm, EtCO <sub>2</sub> alarm, PEEP alarm, Apnea alarm, Battery level alarm, when the parameter monitoring value reaches the alarm condition,the host will generate the above alarm.	<input checked="" type="checkbox"/>	Qualified	<input type="checkbox"/>	Unqualified
Lable and manual	The lable of the T6 should include a) product name,Mode,b)tests per kit ,c)LotNo.,expiration,storage,d)Manufacturer address	<input checked="" type="checkbox"/>	Qualified	<input type="checkbox"/>	Unqualified
Attachment	It should be consistent with the packing list.	<input checked="" type="checkbox"/>	Qualified	<input type="checkbox"/>	Unqualified
package	Package is intact,Lables and indicators are correct.	<input checked="" type="checkbox"/>	Qualified	<input type="checkbox"/>	Unqualified
Carton box	The carton box should be free of stain and impact mark, with flat surface and clear and correct printed label.	<input checked="" type="checkbox"/>	Qualified	<input type="checkbox"/>	Unqualified
conclusion	Qualified				

Checked by/Date: 李赛 2021. 10. 25

Approved by/Date: 王瑞强 2021. 10. 25